

EU DECLARATION OF CONFORMITY

We, the manufacturer, AKSO PRODUTOS ELETRONICOS LTDA, located at Rua Emilio H. Dexheimer, 357 – Bairro Jardim América – CEP: 93032-200 in São Leopoldo – Rio Grande do Sul – Brazil, declare that the autonomous product intended for physical-chemical measurements of water, model

AK90 Tester | Pocket pH Meter 0.0 to 14.0 pH

To which this declaration refers, is in conformity with the provisions of the following Directives:

Directive 2001/95/EC	General Product Safety
Directive 2012/19/EU	Waste electrical and electronic equipment
Directive 2014/35/EU	Low Voltage Directive and its standards
Directive 2014/30/EU	Electromagnetic Compatibility (EMC)

The product complies with the applicable requirements of the following standards:

EN 61000-4-2: 2001	Electromagnetic Compatibility (EMC) - Part 4-2: Tests and measurement techniques – Electrostatic discharge immunity test. 8 kV / 6 kV (air / horizontal and vertical plane)
EN 61000-4-3: 2006	Electromagnetic Compatibility (EMC) - Part 4-3: Test and measurement techniques – Radiated radiofrequency electromagnetic field immunity test. 10 V/m (80 MHz to 3000 MHz)
EN 61000-6-3: 2007	Radiated Disturbances (30MHz to 1000MHz)

I, the undersigned, hereby declare that the aforementioned product is in conformity with the referenced directives, given the instructions contained in the product documentation have been followed.

São Leopoldo, Rio Grande do Sul, Brazil, February, 2026.

Name: Leandro da Silva
Title: Industrial Manager

Signature: 